



KDBEX

EU TYPE EXAMINATION CERTIFICATE

[2] Equipment and protective systems intended for use in potentially explosive atmospheres.
Directive 2014/34/EU (Rozporządzenie Ministra Rozwoju z dnia 06.06.2016 r.
Dz.U. z dnia 09.06.2016 r. Poz. 817)

[3] EU type examination certificate (module B):

KDB 19ATEX0011X

2nd edition

[4] Equipment :

**Smart pressure transmitters type
APC-2000ALW, APC-2000ALW Safety, APC-2000ALW-1;
Smart differential pressure transmitters type
APR-2000ALW, APR-2000ALW Safety, APR-2000ALW/G;
Smart level probes type APR-2000YALW**

[5] Manufacturer:

APLISENS S.A.

[6] Address:

ul. Morelowa 7, 03-192 Warszawa, Poland

[7] The equipment or protective system and any acceptable variations thereto are specified in the schedule to this certificate.

[8] Główny Instytut Górniczo – Państwowy Instytut Badawczy, Notified Body no 1453 according to Directive 2014/34/EU of February 26, 2014, approves that the equipment or protective system specified in this certificate has been found to comply with the essential health and safety requirements for the design and construction of equipment and protective systems intended for use in potentially explosive atmosphere given in Annex II to Directive 2014/34 /EU (Załącznik nr 2 Rozporządzenia Ministra Rozwoju z dnia 06.06.2016 r. Dz.U. z dnia 09.06.2016 r. Poz. 817) . The results of the assessment and examinations as well as the list of agreed documentation are recorded in the confidential Report **KDB No 19.012-2 [T-7517]**

[9] The essential health and safety requirements have been met by compliance with the requirements of the following standards:

**EN IEC 60079-0:2018; EN 60079-1:2014; EN 60079-11:2012;
EN 60079-26:2015; EN 60079-31:2014; EN 50303:2000**

[10] If sign "X" is placed after the certificate number, this means the special conditions of use set out in the schedule to this certificate.

[11] This EU type examination certificate relates only to the construction, assessment and testing of the specified product in accordance with Directive 2014/34/EU (Rozporządzenie Ministra Rozwoju z dnia 06.06.2016 r. Dz.U. z dnia 09.06.2016 r. Poz. 817). The certificate shall not cover the remaining requirements of the Directive regarding the manufacturing process and placing the equipment or protective system on the market.

[12] The marking of the equipment is shown in point [15].

KIEROWNIK
Zespołu ds. Bezpieczeństwa Przeciwwybuchowego
Jednostki Oceny Zgodności
GŁÓWNEGO INSTYTUTU GÓRNICZWA -
Państwowego Instytutu Badawczego

mgr inż. Piotr Madej

ATEX Certification
Expert



KIEROWNIK
Jednostki Oceny Zgodności
Głównego Instytutu Górniczo –
Państwowego Instytutu Badawczego
dr inż. Dariusz Stefaniak

Date of issue: **26 January 2026**

Page 1 of 5

[13]

[14]

SCHEDULE
EU type examination certificate
KDB 19ATEX0011X 2nd edition



[15] Description:

Pressure transmitters type APC-2000ALW, APC-2000ALW Safety, APC-2000ALW-1; differential pressure transmitters type APR-2000ALW, APR-2000ALW Safety, APR-2000ALW/G and level probes type APR-2000YALW convert resistance changes proportional to the measured pressure of piezoresistive bridge, located in the single crystal of silicon diaphragm, into a standard current signal $4 \div 20$ mA with HART communications signal. Transmitters can be used for measurement of dense and aggressive media, at high and low temperatures.

The basic unit of the transmitter and the probe is a measuring head with a silicon diaphragm sensor, working in the intrinsically safe circuit (Ex ia), mounted in transmitter enclosure. Measuring heads can be equipped with different pressure connections. Inside the head there is the "pressure chamber" filled with manometer liquid. It is limited by a diaphragm welded tightly to the head's body, on the side of measured medium. Differential pressure transmitters have two separated diaphragms for the inputs: "+" and "-". Inside the head there is a bushing in which a measuring silicon diaphragm with piezoresistors is installed. In pressure and differential pressure transmitters it is allowed to cover diaphragm seals with PTFE foil.

Enclosures of transmitters are made of die-cast aluminium alloy or stainless steel. Enclosure consists of a body and two screwed covers (display cover and electrical connection cover). The cable is introduced into the enclosure by cable gland with thread M20x1,5 or 1/2NPT depending on the version of the enclosure body. In the non-used opening there is mounted plug (cap).

The device version including the flameproof enclosure requires use of flameproof cable gland and plug. Cable entries and plugs should also meet the requirements for covers protected by the enclosure (Ex tb). The device in the flameproof and dustproof version includes plug produced by Aplisens S.A..

The measuring head working in the intrinsically safe circuit (Ex ia), in the version of the device including the flameproof enclosure, is separated from the rest of the equipment by the bushing.

In the transmitter enclosure, is also installed a terminal strip allowing additional connection of the communicator and measurement of the output current, without interrupting the circuit.




SCHEDULE
EU type examination certificate
KDB 19ATEX0011X 2nd edition





Marking:


I M2 Ex db ia I Mb *


II 1/2G Ex ia/db IIC T6/T5 Ga/Gb

 II 1/2D Ex ia/tb IIIC T105°C Da/Db
or
I M2 Ex db ia I Mb *

 II 2G Ex db ia IIC T6/T5 Gb
II 2D Ex ia tb IIIC T105°C Db
or
I M1 Ex ia I Ma *

 II 1/2G Ex ia IIC T5/T4 Ga/Gb
II 1D Ex ia IIIC T105°C Da
or

 II 1/2G Ex ia/db IIC T6/T5 Ga/Gb
or

 II 1/2G Ex ia IIC T5/T4 Ga/Gb
* - only the version in a steel enclosure

Technical parameters:

Output signal:

4 ÷ 20mA in a two-wire system + HART

Device version Ex ia/db, Ex ia/tb, Ex db ia, Ex ia tb:

Supply voltage:

U_{max}=55V DC, U_m=250V AC (APC-2000ALW, APR-2000ALW, APR-2000ALW/G, APR-2000YALW)

U_{max}=36V DC, U_m=250V AC (APC-2000ALW Safety, APR-2000ALW Safety, APC-2000ALW-1)

Ambient temperature: -40°C ÷ 40°C (special version: from -50°C)

Temperature class: T6

Ambient temperature: -40°C ÷ 75°C (special version: from -50°C)

Temperature class: T5

Ambient temperature: -25°C ÷ 40°C - type APC-2000ALW-1

Temperature class: T6

Ambient temperature: -25°C ÷ 55°C - type APC-2000ALW-1

Temperature class: T5

Maximum surface temperature

- version Ex ia/tb: 105°C

Degree of protection: IP66 / IP67

Główny Instytut Górnictwa – Państwowy Instytut Badawczy, 40-166 Katowice, Plac Gwarków 1
Jednostka Oceny Zgodności, 43-190 Mikołów, ul. Podleska 72

This certificate may only be reproduced in its entirety.





Device version Ex ia:

Supply voltage:	U _{max} = 30V DC
Ambient temperature:	-40 ÷ 80°C (special version: from -50°C)
Temperature class:	T5/T4
Ambient temperature:	-25°C ÷ 55°C - type APC-2000ALW-1
Temperature class:	T5/T4
Maximum surface temperature:	105°C
Degree of protection:	IP66 / IP67

Intrinsically safe parameters:

Supply from a power source with linear output characteristic:

U _i =30V	Li = 18μH	Temperature class: T5
I _i =100mA	Ci = 2,5nF*	
Pi=0,75W		

Supply from a power source with trapezoidal output characteristic:

U _i =24V	Li = 18μH	Temperature class: T5
I _i =50mA	Ci = 2,5nF*	
Pi=0,7W		

Supply from a power source with rectangular output characteristic:

U _i =24V	Li = 18μH	Temperature class: T5
I _i =25mA	Ci = 2,5nF*	
Pi=0,6W		

U _i =24V	Li = 18μH	Temperature class: T4
I _i =50mA	Ci = 2,5nF*	
Pi=1,2W		


*) 7,5nF for type APC-2000ALW-1

[16] Test Report:

"ATEX assessment report" KDB No 19.012-2

[17] Special conditions of use:

- The maximum temperature of the external heating source cannot heat the transmitter above the maximum declared ambient temperature.
- The flameproof joints are not intended to be repaired.
- In hazardous zones of dust explosion, transmitters with painted enclosures, as well as transmitters equipped with plastic marking plates and diaphragm separator elements covered with PTFE layer, should be installed in a way that prevents electrostatic charging, in accordance with the instructions.



SCHEDULE
EU type examination certificate
KDB 19ATEX0011X 2nd edition



- In hazardous zones of explosion, diaphragm separators covered PTFE layer, should be installed in places and in a way that prevents electrostatic charging.
- The diaphragm separator containing titanium elements must be protected against mechanical impacts.
- An intrinsically safe transmitter version with surge arrester, marked on the nameplate as "SA" does not meet the requirements of clause 10.3 of EN 60079-11 (500Vrms). The device should be installed in accordance with the instructions.
- In the case of use a transmitter with a nameplate containing various types of explosion-proof execution, the type of protection must be permanently marked on the nameplate before installation, according to the instructions.
- In the device version including the flameproof enclosure, the diaphragm should not be subject on damage during installation and exploitation of the transmitter. The transmitter diaphragm is made of stainless steel, Hastelloy alloy or tantalum and must not be exposed to medium that could cause its damage.
- Smart pressure transmitters type APC-2000ALW-1 in Ex ia/db and Ex ia versions, as well as other Ex ia pressure transmitters with PM12 connection, are approved for use exclusively in gas hazardous areas (Group II).

[18] Essential health and safety requirements:

Met by fulfilling the requirements of the following standards:

EN IEC 60079-0:2018	(PN-EN IEC 60079-0:2018-09)
EN 60079-1:2014	(PN-EN 60079-1:2014-12)
EN 60079-11:2012	(PN-EN 60079-11:2012)
EN 60079-26:2015	(PN-EN 60079-26:2015-04)
EN 60079-31:2014	(PN-EN 60079-31:2014-10)
EN 50303:2000	(PN-EN 50303:2004)

Document history:

- EU type examination certificate KDB 19ATEX0011X, 0 edition of 30.04.2019, initial certification.
- EU type examination certificate KDB 19ATEX0011X, 1st edition of 28.10.2022 replaces EU type examination certificate KDB 19ATEX0011X, 0 edition of 30.04.2019.
Type names for differential pressure transmitters have been changed. New versions of the MPC-FH-Exi-Exd-rev2 and MPC5-rev2.1.2 board assembly have been added. Seals made of NBR material have also been introduced for the Exd and Ext versions. New versions of Exd bushings have been added. The marking of the explosion-proof version has been changed. A special conditions of use has been changed.
- EU type examination certificate KDB 19ATEX0011X, 2nd edition of 26.01.2026 replaces EU type examination certificate KDB 19ATEX0011X, 1st edition of 28.10.2022.
A new type name for the pressure transducer has been added: APC-2000ALW-1. New versions with a connection board (filter) have been added. PM12 connector has been added.

**Główny Instytut Górniczo – Państwowy Instytut Badawczy, 40-166 Katowice, Plac Gwarków 1
Jednostka Oceny Zgodności, 43-190 Mikołów, ul. Podleska 72**

This certificate may only be reproduced in its entirety.

